proval Listing Number - s.WA-973 ssued in advance of preparation of draw Approval Listing only in case of eqrim Approval).

ny Requesting Approval -

Ross Cunningham Industries Ltd.

4600 Burgoyne St.

Mississauga, Ontario L4W 1Gl

pe of Device Meter for lubricating oils, mfd. by Grover Mfg. Corp.

Montebello, California, U.S.A.

del No.

Rating or Capacity

Remarks

1584-P

 $\frac{1}{2}$ - 3 gpm

Meters for all grades of oil to be factory tested with ASAE30 oil.

Grover meters use the same oscillating piston type of metering element as is used in Badger meters for automative lubricating oils.

All meters to comply with Spec. S-VM-2A

te of Application for Approval - April 22, 1976

le, No. 6953-1

Dec 9, 1976

ate of Filing

Signature

TE: File Record of Approval Action with Approval Listing notices when roval Listing number is assigned in advance of preparation of notice; file at k of book by Central File No. when Approval Listing number is not assigned.

re Record of Approval Action from Approval Listing Book when the company inotified of rejection, or when Special Approval Listing notice or draft of approval Listing notice is filed with notices; replace draft of Approval Listing otice with printed notice when available.