



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario

Avenue Holland
Ottawa, Ontario

Legal Metrology Laboratories

Telephone: (613) 952-0631

Your file Votre référence

Our file Notre référence

06953-AP-VL-88-0027

AP-VL-88-0027

August 5, 1988

AML-V-27

Specialty Process Systems Ltd.
82 King Street
Georgetown, Ontario
L7G 2G7
Attention: Mr. G. Collins

Dear Mr. Collins:


This is in response to your letter dated July 26, 1988, regarding Notices of Approval S.WA-2145, and proposing to modify the models 10-3325 and 10-3327 approved therein by relocating the filter from the inlet of the pump to the inlet of the meter.

I am advised that, upon evaluation of the proposed modifications, the Volumetric Laboratory has concluded that they may be regarded as immaterial in relation to the approval provisions of the Weights and Measures Act. Accordingly, pursuant to the authority delegated to me under the said Act:

1. your letter is hereby acknowledged as satisfying the notification requirements of section 14, of the Weights and Measures Regulations, SOR/85-736, and
2. the proposed alteration is hereby accepted, in accordance with the aforementioned section, as being an immaterial one and hence included within the approval.

All recipients of Notices of Approval under the Weights and Measures Act are being informed of this decision by copy of this letter.

Yours truly,


W.R. Virtue
Chief

cc: Mailing List (Volumetric Approvals)

Canada

**The self-contained unit
consists of:**

1. Air eliminator / Dés aérateur
2. Filter / Filtre
3. Centrifugal pump / Pompe centri fuge
4. Control valve / Clapet de contrôle
5. OT liquid meter/ Compteur liqui de
6. Check valve / Clapet de retenue
7. Control unit / Regis tre de contrôle

