



Date: August 27, 2009

File number: AP-AV-09-0022

**MODIFICATION ACCEPTANCE LETTER (MAL)
FOR
Approved Measuring Devices**

Type of Device

Bulk Liquid Meter

Purpose

The purpose of this letter is to convey details of modifications to liquid measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect devices identified in the following Notice(s) of Approval:

Approval Number

SWA-0368

Applicant

Liquid Controls
A unit of IDEX Corporation
105 Albrecht Drive
Lake Bluff, Illinois, 60044-2242

Description of Modifications

-The density ranges as listed in approval SWA-0368 are superseded by density ranges as listed in bulletin V-16.

Original copy signed by:

Amina Abid
Manager-Liquids Measurement



Date: 27 Août, 2009

Dossier: AP-AV-09-0022

LETTRE D'ACCEPTATION DE MODIFICATION (LAM)
pour
Compteurs et appareils approuvés

Type d'appareil

Compteur pour liquids en vrac

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux appareils visés par l'avis d'approbation suivant:

Numéro d'approbation

SWA-0368

Requérant

Liquid Controls
A unit of IDEX Corporation
105 Albrecht Drive
Lake Bluff, Illinois, 60044-2242

Description des modifications

Les plages de masse volumique tel qu'énumérées dans l'approbation SWA-0368 sont remplacées par les plages de masse volumique énumérées dans le bulletin V-16.

Copie authentique signée par :

Amina Abid
Gestionnaire de mesure des liquides